
Midmark M11 - v1.0.0
Total Cycles: 2082

Name: RITTER 1
Sterilizer ID: V2178198
07/04/2022 12:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.2 F 0.0
2:00 187.5 F 3.5
4:00 203.7 F 7.5
6:00 221.8 F 11.9
8:00 221.3 F 5.9
10:00 234.1 F 10.8
12:00 230.9 F 6.9
14:00 244.0 F 13.1
16:00 242.3 F 11.3
18:00 254.7 F 17.7
20:00 265.7 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.5 F 28.9
2:00 271.4 F 28.2
3:00 272.5 F 29.0
4:00 271.9 F 28.4

Min 271.4 F 28.1
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.8 F 0.0
5:00 218.6 F 0.0
10:00 233.2 F 0.0
15:00 231.2 F 0.0
20:00 229.5 F 0.0
25:00 230.7 F 0.0
30:00 228.1 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 21:45
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:58
CYCLE COMPLETE

07/04/2022 01:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____