
Midmark M11 - v1.0.5
Total Cycles: 31

Name: RITTER2
Sterilizer ID: V2313228
05/19/2021 11:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.7 F 0.0
2:00 184.7 F 3.8
4:00 203.0 F 9.1
6:00 221.3 F 7.6
8:00 231.6 F 10.8
10:00 234.1 F 8.7
12:00 239.6 F 10.2
14:00 248.7 F 14.9
16:00 261.0 F 21.9
18:00 272.1 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.2 F 29.1
2:00 272.3 F 29.6
3:00 271.7 F 28.7
4:00 272.7 F 29.5

Min 271.5 F 28.7
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.7 F 0.0
5:00 230.5 F 0.0
10:00 226.7 F 0.0
15:00 223.6 F 0.0
20:00 220.5 F 0.0
25:00 218.5 F 0.0
30:00 214.0 F 0.0
Drying Complete

FILLING: 2:12
HEATING: 18:27
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 00:57:43
CYCLE COMPLETE

05/19/2021 12:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____