

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000874  
16:28 17/01/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:42  
PRESSURIZING 3:00  
STERILIZING 4:19  
136.0 C 227kPa 4:19  
136.3 C 231kPa 4:49  
136.3 C 231kPa 5:19  
136.4 C 233kPa 5:49  
136.2 C 231kPa 6:19  
136.3 C 231kPa 6:49  
136.4 C 233kPa 7:19  
136.4 C 233kPa 7:49  
136.4 C 233kPa 7:49  
VENTING 7:49  
AIR DRYING 8:46  
CYCLE COMPLETE 14:34

-----  
Digital Signature #  
9B35FCE93FADEF9C  
-----