
Midmark M11 - v1.0.5
Total Cycles: 1903

Name: RITTER2
Sterilizer ID: V2313228
01/05/2023 04:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.0 F 0.0
2:00 156.5 F 3.0
4:00 176.5 F 7.3
6:00 199.9 F 12.8
8:00 219.9 F 6.8
10:00 234.0 F 12.3
12:00 236.4 F 10.0
14:00 236.1 F 8.6
16:00 250.4 F 15.9
18:00 262.9 F 23.0
20:00 272.0 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.4 F 29.2
2:00 272.5 F 29.7
3:00 271.9 F 28.8
4:00 272.8 F 29.6

Min 271.6 F 28.7
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.5 F 0.0
5:00 223.4 F 0.0
10:00 228.1 F 0.0
15:00 233.8 F 0.0
20:00 223.0 F 0.0
25:00 221.2 F 0.0
30:00 218.9 F 0.0
Drying Complete

FILLING: 1:24
HEATING: 20:07
STERILIZING: 4:00
VENTING: 2:56
DRYING: 30:00
TOTAL CYCLE: 00:58:27
CYCLE COMPLETE

01/05/2023 05:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____