
Midmark M11 - v1.0.5
Total Cycles: 3794

Name: RITTER2
Sterilizer ID: V2313228
07/17/2024 01:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 86.4 F 0.0
2:00 86.8 F 0.0
4:00 91.5 F 0.1
6:00 132.9 F 0.3
8:00 157.1 F 2.4
10:00 175.7 F 5.4
12:00 200.5 F 8.7
14:00 213.5 F 12.3
16:00 220.5 F 6.6
18:00 231.8 F 11.1
20:00 231.2 F 7.9
22:00 243.2 F 13.5
24:00 242.0 F 12.0
26:00 253.1 F 17.8
28:00 263.3 F 23.7
30:00 272.7 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.4
1:00 273.0 F 30.0
2:00 271.6 F 29.3
3:00 272.0 F 30.0
4:00 272.4 F 29.8

Min 271.1 F 29.2
Max 273.0 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.3 F 0.0
5:00 217.7 F 0.0
10:00 227.5 F 0.0
15:00 213.3 F 0.0
20:00 215.9 F 0.0
25:00 219.9 F 0.0
30:00 223.7 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 30:21
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:09:01
CYCLE COMPLETE

07/17/2024 03:04 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____