
Midmark M11 - v1.0.0
Total Cycles: 2934

Name: RITTER 1
Sterilizer ID: V2178198
02/22/2023 05:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.9 F 0.0
2:00 189.3 F 4.6
4:00 208.2 F 9.7
6:00 223.6 F 5.8
8:00 235.9 F 11.1
10:00 234.7 F 8.6
12:00 240.6 F 9.8
14:00 248.2 F 14.3
16:00 261.5 F 21.5
18:00 272.7 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.0 F 28.4
2:00 272.4 F 28.8
3:00 271.5 F 28.6
4:00 271.9 F 28.4

Min 271.3 F 28.2
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.6 F 0.0
5:00 230.5 F 0.0
10:00 242.8 F 0.0
15:00 239.1 F 0.0
20:00 232.8 F 0.0
25:00 225.2 F 0.0
30:00 223.5 F 0.0
Drying Complete

FILLING: 2:19
HEATING: 18:19
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:57:55
CYCLE COMPLETE

02/22/2023 06:38 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____