
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001097
10:41 26/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:18
PRESSURIZING 3:29
STERILIZING 5:03
136.4 C 223kPa 5:03
136.3 C 225kPa 5:33
136.0 C 222kPa 6:03
136.4 C 224kPa 6:33
136.2 C 222kPa 7:03
136.4 C 225kPa 7:33
136.4 C 225kPa 8:03
136.3 C 223kPa 8:33
136.3 C 223kPa 8:33
VENTING 8:33
AIR DRYING 9:32
CYCLE COMPLETE 9:32

Digital Signature #
CA083BFA4E855BC1
