
Midmark M11 - v1.0.5
Total Cycles: 782

Name: RITTER2
Sterilizer ID: V2313228
02/10/2022 01:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.4 F 0.0
2:00 187.0 F 4.3
4:00 207.3 F 10.1
6:00 224.3 F 6.8
8:00 236.2 F 12.7
10:00 238.4 F 11.2
12:00 239.1 F 10.6
14:00 253.5 F 17.9
16:00 265.6 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.1
1:00 272.4 F 30.1
2:00 272.1 F 29.4
3:00 271.1 F 29.0
4:00 271.7 F 29.8

Min 271.0 F 29.0
Max 273.1 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.2 F 0.0
5:00 235.7 F 0.0
10:00 221.4 F 0.0
15:00 222.0 F 0.0
20:00 172.3 F 0.0
25:00 168.4 F 0.0
30:00 171.7 F 0.0
Drying Complete

FILLING: 2:41
HEATING: 17:39
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:57:35
CYCLE COMPLETE

02/10/2022 02:18 PM
ACCEPTED: X
REJECTED:
Operator: U02