

-----  
Midmark M11 - v1.0.5  
Total Cycles: 626

Name: RITTER2  
Sterilizer ID: V2313228  
12/13/2021 09:44 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 111.0 F 0.0  
2:00 113.4 F 0.5  
4:00 129.1 F 2.5  
6:00 158.6 F 5.7  
8:00 182.6 F 10.0  
10:00 209.1 F 6.0  
12:00 217.4 F 7.5  
14:00 229.8 F 12.3  
16:00 233.5 F 9.3  
18:00 238.2 F 9.9  
20:00 245.9 F 13.9  
22:00 258.0 F 20.4  
24:00 268.6 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.3  
1:00 272.6 F 30.2  
2:00 272.3 F 29.6  
3:00 271.2 F 29.4  
4:00 272.3 F 30.2

Min 271.2 F 29.2  
Max 272.9 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.8 F 0.0  
5:00 202.7 F 0.0  
10:00 209.9 F 0.0  
15:00 216.2 F 0.0  
20:00 214.0 F 0.0  
25:00 215.0 F 0.0  
30:00 215.5 F 0.0  
Drying Complete

FILLING: 1:11  
HEATING: 25:17  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:03:37  
CYCLE COMPLETE

12/13/2021 10:47 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_