
Midmark M11 - v1.0.0
Total Cycles: 737

Name: RITTER 1
Sterilizer ID: V2178198
03/22/2021 01:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 99.6 F 0.0
2:00 130.5 F 0.0
4:00 145.7 F 0.1
6:00 189.6 F 0.3
8:00 192.5 F 4.3
10:00 207.9 F 8.2
12:00 221.3 F 12.5
14:00 232.4 F 7.8
16:00 239.7 F 13.1
18:00 240.8 F 10.4
20:00 235.5 F 8.4
22:00 251.2 F 15.5
24:00 263.4 F 22.5
26:00 270.7 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 271.5 F 28.4
2:00 271.5 F 28.5
3:00 271.4 F 28.5
4:00 271.6 F 28.4

Min 271.3 F 28.3
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.1 F 0.0
5:00 231.7 F 0.0
10:00 221.1 F 0.0
15:00 227.7 F 0.0
20:00 229.7 F 0.0
25:00 197.8 F 0.0
30:00 200.7 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 26:37
STERILIZING: 4:00
VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 01:04:27
CYCLE COMPLETE

03/22/2021 02:23 PM
ACCEPTED: X
REJECTED:
Operator: U01