
Midmark M11 - v1.0.0
Total Cycles: 4034

Name: RITTER 1
Sterilizer ID: V2178198
11/17/2023 11:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 129.8 F 0.0
2:00 128.7 F 1.2
4:00 135.5 F 3.5
6:00 165.5 F 6.5
8:00 187.3 F 10.6
10:00 214.0 F 4.6
12:00 214.7 F 6.7
14:00 225.3 F 10.7
16:00 230.2 F 7.2
18:00 238.9 F 11.8
20:00 237.7 F 9.8
22:00 250.1 F 15.7
24:00 260.9 F 21.8
26:00 271.2 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.5 F 29.3
2:00 271.7 F 28.7
3:00 271.5 F 29.0
4:00 272.4 F 29.3

Min 271.2 F 28.5
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.6 F 0.0
5:00 213.5 F 0.0
10:00 216.8 F 0.0
15:00 215.5 F 0.0
20:00 219.8 F 0.0
25:00 234.1 F 0.0
30:00 228.3 F 0.0

Drying Complete

FILLING: 1:10
HEATING: 26:45
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:05:36
CYCLE COMPLETE

11/17/2023 12:08 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____