
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002076
8:44 20/07/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:52
PRESSURIZING 3:53
STERILIZING 4:56
136.4 C 221kPa 4:56
136.3 C 223kPa 5:26
136.4 C 221kPa 5:56
136.2 C 222kPa 6:26
136.2 C 220kPa 6:56
136.4 C 224kPa 7:26
136.2 C 222kPa 7:56
136.4 C 224kPa 8:26
136.4 C 224kPa 8:26
VENTING 8:26
AIR DRYING 9:21
CYCLE COMPLETE 9:25

Digital Signature #
A0CE982A61A70625
