
Midmark M11 - v1.0.5
Total Cycles: 3979

Name: RITTER2
Sterilizer ID: V2313228
09/27/2024 10:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.4 F 0.0
2:00 76.4 F 0.0
4:00 82.6 F 0.0
6:00 113.7 F 0.1
8:00 154.8 F 0.4
10:00 183.7 F 0.7
12:00 203.9 F 3.3
14:00 215.1 F 3.9
16:00 226.0 F 7.5
18:00 236.1 F 11.5
20:00 232.9 F 7.9
22:00 245.4 F 13.0
24:00 243.7 F 12.2
26:00 241.7 F 11.2
28:00 255.5 F 18.3
30:00 267.2 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.6 F 28.8
2:00 272.6 F 29.3
3:00 271.8 F 28.3
4:00 273.0 F 29.2

Min 271.6 F 28.3
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.9 F 0.0
5:00 182.7 F 0.0
10:00 204.9 F 0.0
15:00 201.9 F 0.0
20:00 197.3 F 0.0
25:00 199.4 F 0.0
30:00 203.8 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 31:24
STERILIZING: 4:00
VENTING: 2:60
DRYING: 30:00
TOTAL CYCLE: 01:09:17
CYCLE COMPLETE

09/27/2024 11:34 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____