
Midmark M11 - v1.0.5
Total Cycles: 2028

Name: RITTER2
Sterilizer ID: V2313228
02/09/2023 08:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 77.5 F 0.0
2:00 92.6 F 0.0
4:00 84.2 F 0.0
6:00 109.4 F 0.0
8:00 147.0 F 0.1
10:00 181.8 F 2.9
12:00 210.4 F 5.9
14:00 221.0 F 9.8
16:00 228.2 F 10.2
18:00 232.1 F 9.3
20:00 240.0 F 10.8
22:00 242.5 F 12.2
24:00 241.0 F 10.9
26:00 253.6 F 17.4
28:00 264.9 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.3 F 29.1
2:00 271.4 F 28.9
3:00 272.2 F 29.1
4:00 271.2 F 28.7

Min 271.2 F 28.5
Max 272.5 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.8 F 0.0
5:00 217.8 F 0.0
10:00 233.8 F 0.0
15:00 230.7 F 0.0
20:00 230.6 F 0.0
25:00 230.9 F 0.0
30:00 210.8 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 29:53
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:07:49
CYCLE COMPLETE

02/09/2023 09:47 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____