
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 000602
10:21 26/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:13
CYCLE ABORTED 3:19
WATER LEVEL LOW
NOT STERILE
END TIME 3:19

Digital Signature #
2277F94CE13E0832
