

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001200  
10:03 29/07/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:20  
PRESSURIZING 3:25  
STERILIZING 4:37  
136.5 C 224kPa 4:37  
136.4 C 224kPa 5:07  
136.4 C 225kPa 5:37  
136.4 C 224kPa 6:07  
136.4 C 224kPa 6:37  
136.1 C 222kPa 7:07  
136.2 C 222kPa 7:37  
136.1 C 222kPa 8:07  
136.1 C 222kPa 8:07  
VENTING 8:07  
AIR DRYING 9:03  
CYCLE COMPLETE 24:03  
-----

Digital Signature #  
B348DF79AE5AEA05  
-----