
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001492
15:02 26/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:09
PRESSURIZING 3:12
STERILIZING 4:12
136.4 C 223kPa 4:12
136.3 C 224kPa 4:42
136.2 C 222kPa 5:12
136.3 C 224kPa 5:42
136.4 C 226kPa 6:12
136.3 C 224kPa 6:42
136.4 C 224kPa 7:12
136.4 C 225kPa 7:42
136.4 C 225kPa 7:42
VENTING 7:42
AIR DRYING 8:37
CYCLE COMPLETE 8:41

Digital Signature #
2C999F018EBCD2B4
