

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1792

Name: RITTER2  
Sterilizer ID: V2313228  
11/30/2022 11:47 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 197.5 F 0.0  
2:00 189.9 F 2.7  
4:00 206.9 F 8.4  
6:00 223.8 F 10.0  
8:00 235.0 F 11.2  
10:00 236.4 F 9.8  
12:00 237.5 F 9.0  
14:00 251.3 F 16.6  
16:00 264.0 F 24.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.2  
1:00 272.4 F 29.6  
2:00 271.8 F 29.7  
3:00 272.2 F 29.5  
4:00 271.7 F 29.8  
  
Min 271.4 F 29.1  
Max 272.7 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.2 F 0.0  
5:00 224.6 F 0.0  
10:00 232.5 F 0.0  
15:00 225.8 F 0.0  
20:00 222.6 F 0.0  
25:00 220.6 F 0.0  
30:00 219.3 F 0.0  
Drying Complete

FILLING: 1:34  
HEATING: 17:52  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:56:40  
CYCLE COMPLETE

11/30/2022 12:43 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_