
Midmark M11 - v1.0.5
Total Cycles: 2304

Name: RITTER2
Sterilizer ID: V2313228
04/27/2023 09:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 146.6 F 0.0
2:00 146.1 F 0.0
4:00 154.5 F 1.6
6:00 180.3 F 5.6
8:00 199.4 F 9.9
10:00 215.9 F 8.0
12:00 225.9 F 8.1
14:00 237.0 F 12.8
16:00 237.6 F 10.1
18:00 237.4 F 9.1
20:00 248.9 F 14.9
22:00 260.1 F 21.2
24:00 270.6 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 272.4 F 29.2
2:00 271.9 F 29.2
3:00 272.2 F 29.1
4:00 271.7 F 29.1

Min 271.5 F 28.7
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.2 F 0.0
5:00 210.4 F 0.0
10:00 208.0 F 0.0
15:00 197.5 F 0.0
20:00 194.1 F 0.0
25:00 191.5 F 0.0
30:00 188.7 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 24:49
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:03:09
CYCLE COMPLETE

04/27/2023 10:07 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____