
Midmark M11 - v1.0.5
Total Cycles: 3420

Name: RITTER2
Sterilizer ID: V2313228
02/29/2024 09:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.7 F 0.0
2:00 70.5 F 0.0
4:00 75.4 F 0.0
6:00 105.2 F 1.0
8:00 143.8 F 3.8
10:00 169.7 F 6.8
12:00 188.6 F 10.7
14:00 215.3 F 5.3
16:00 218.5 F 7.5
18:00 229.8 F 11.5
20:00 230.0 F 7.3
22:00 242.2 F 12.6
24:00 239.9 F 10.7
26:00 252.4 F 16.9
28:00 263.3 F 23.4
30:00 272.6 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.8
1:00 272.4 F 29.6
2:00 271.8 F 28.7
3:00 271.5 F 29.3
4:00 272.4 F 29.4

Min 271.0 F 28.6
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.3 F 0.0
5:00 235.0 F 0.0
10:00 205.7 F 0.0
15:00 191.7 F 0.0
20:00 183.7 F 0.0
25:00 185.9 F 0.0
30:00 186.7 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:16
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:08:18
CYCLE COMPLETE

02/29/2024 10:15 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____