
Midmark M11 - v1.0.0
Total Cycles: 3036

Name: RITTER 1
Sterilizer ID: V2178198
03/27/2023 10:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 146.4 F 0.0
2:00 147.5 F 2.7
4:00 165.4 F 6.0
6:00 187.1 F 10.2
8:00 206.9 F 5.5
10:00 219.1 F 8.3
12:00 232.1 F 13.1
14:00 234.8 F 9.6
16:00 237.5 F 9.1
18:00 246.4 F 13.7
20:00 258.4 F 20.2
22:00 269.4 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.7 F 29.3
2:00 271.6 F 28.5
3:00 272.6 F 29.4
4:00 272.2 F 28.9

Min 271.5 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.7 F 0.0
5:00 213.2 F 0.0
10:00 228.0 F 0.0
15:00 230.6 F 0.0
20:00 225.4 F 0.0
25:00 224.5 F 0.0
30:00 224.8 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 23:02
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:01:13
CYCLE COMPLETE

03/27/2023 11:21 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____