
Midmark M11 - v1.0.0
Total Cycles: 3777

Name: RITTER 1
Sterilizer ID: V2178198
09/14/2023 04:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.0 F 0.0
2:00 176.3 F 0.3
4:00 190.2 F 4.0
6:00 208.2 F 8.1
8:00 224.4 F 13.4
10:00 232.2 F 9.0
12:00 237.5 F 10.6
14:00 242.6 F 12.0
16:00 240.8 F 10.7
18:00 253.8 F 17.0
20:00 264.9 F 23.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.0
1:00 272.1 F 28.6
2:00 271.2 F 27.9
3:00 272.3 F 28.8
4:00 271.6 F 28.2

Min 271.1 F 27.9
Max 272.4 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.2 F 0.0
5:00 222.8 F 0.0
10:00 242.3 F 0.0
15:00 233.2 F 0.0
20:00 221.4 F 0.0
25:00 217.5 F 0.0
30:00 219.0 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 21:55
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:00:43
CYCLE COMPLETE

09/14/2023 05:56 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____