

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000244  
11:35 29/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:27  
PRESSURIZING 3:32  
STERILIZING 5:09  
136.4 C 223kPa 5:09  
136.4 C 233kPa 5:39  
136.4 C 233kPa 6:09  
136.2 C 232kPa 6:39  
136.2 C 232kPa 7:09  
136.1 C 232kPa 7:39  
136.3 C 233kPa 8:09  
136.4 C 233kPa 8:39  
136.4 C 233kPa 8:39  
VENTING 8:39  
AIR DRYING 9:41  
CYCLE COMPLETE 16:32

-----  
Digital Signature #  
E2945EC44F225466  
-----