
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007521
8:05 10/06/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:23
PRESSURIZING 2:26
STERILIZING 3:25
136.4 C 220kPa 3:25
136.0 C 221kPa 3:55
136.4 C 223kPa 4:25
136.4 C 224kPa 4:55
136.2 C 221kPa 5:25
136.2 C 223kPa 5:55
136.1 C 221kPa 6:25
136.3 C 219kPa 6:55
136.3 C 219kPa 6:55
VENTING 6:55
AIR DRYING 7:45
CYCLE COMPLETE 8:01

Digital Signature #
56654D63AF0B919F
