
Midmark M11 - v1.0.0
Total Cycles: 5276

Name: RITTER 1
Sterilizer ID: V2178198
01/24/2025 04:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 120.1 F 0.0
2:00 123.8 F 1.8
4:00 151.6 F 5.2
6:00 175.5 F 8.7
8:00 196.5 F 12.6
10:00 215.0 F 4.8
12:00 222.3 F 9.1
14:00 234.2 F 13.5
16:00 236.1 F 9.1
18:00 242.0 F 10.9
20:00 244.3 F 12.1
22:00 255.5 F 18.1
24:00 265.8 F 24.3
26:00 271.8 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.7 F 29.0
2:00 271.7 F 28.6
3:00 272.5 F 28.9
4:00 271.6 F 28.4

Min 271.5 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.8 F 0.0
5:00 212.3 F 0.0
10:00 196.8 F 0.0
15:00 195.8 F 0.0
20:00 203.5 F 0.0
25:00 234.2 F 0.0
30:00 227.4 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 26:34
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:05:11
CYCLE COMPLETE

01/24/2025 05:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____