

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2407

Name: RITTER2  
Sterilizer ID: V2313228  
05/24/2023 11:14 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 181.9 F 0.0  
2:00 178.4 F 3.3  
4:00 183.3 F 7.8  
6:00 204.8 F 13.1  
8:00 225.5 F 8.2  
10:00 237.2 F 12.6  
12:00 239.4 F 11.4  
14:00 238.8 F 10.2  
16:00 252.8 F 17.2  
18:00 264.5 F 24.2  
20:00 270.8 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.4  
1:00 272.1 F 29.6  
2:00 272.2 F 29.8  
3:00 272.5 F 29.8  
4:00 272.6 F 29.8

Min 271.7 F 29.2  
Max 272.9 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.2 F 0.0  
5:00 217.2 F 0.0  
10:00 239.1 F 0.0  
15:00 239.0 F 0.0  
20:00 228.9 F 0.0  
25:00 228.5 F 0.0  
30:00 227.5 F 0.0  
Drying Complete

FILLING: 1:29  
HEATING: 20:36  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:59:22  
CYCLE COMPLETE

05/24/2023 12:13 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_