
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001332
14:45 03/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:41
PRESSURIZING 4:41
STERILIZING 6:07
136.0 C 221kPa 6:07
136.4 C 222kPa 6:37
136.3 C 223kPa 7:07
136.3 C 223kPa 7:37
136.4 C 225kPa 8:07
136.3 C 222kPa 8:37
136.5 C 225kPa 9:07
136.4 C 224kPa 9:37
136.4 C 224kPa 9:37
VENTING 9:37
AIR DRYING 10:38
CYCLE COMPLETE 25:38

Digital Signature #
85DCDD86674334A7
