
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000074
11:01 07/08/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:41
PRESSURIZING 3:42
STERILIZING 4:58
136.2 C 222kPa 4:58
136.2 C 223kPa 5:28
136.4 C 222kPa 5:58
136.0 C 222kPa 6:28
136.4 C 224kPa 6:58
136.1 C 222kPa 7:28
136.4 C 224kPa 7:58
135.9 C 222kPa 8:28
135.9 C 222kPa 8:28
VENTING 8:28
AIR DRYING 9:23
CYCLE COMPLETE 24:23

Digital Signature #
4665B2F1B5329906
