
Midmark M11 - v1.0.0
Total Cycles: 2557

Name: RITTER 1
Sterilizer ID: V2178198
11/15/2022 08:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 79.3 F 0.0
2:00 79.7 F 0.0
4:00 92.4 F 0.1
6:00 118.4 F 1.8
8:00 146.1 F 4.6
10:00 176.7 F 8.0
12:00 194.9 F 11.4
14:00 213.7 F 3.7
16:00 223.6 F 8.0
18:00 236.3 F 12.1
20:00 230.6 F 7.7
22:00 244.3 F 12.9
24:00 241.1 F 10.5
26:00 253.3 F 16.6
28:00 263.9 F 22.9
30:00 271.1 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.8 F 29.0
2:00 271.7 F 28.8
3:00 272.7 F 28.9
4:00 271.8 F 28.9

Min 271.3 F 28.4
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 170.7 F 0.0
5:00 214.0 F 0.0
10:00 195.0 F 0.0
15:00 183.3 F 0.0
20:00 180.4 F 0.0
25:00 177.7 F 0.0
30:00 174.9 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 30:52
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:08:56
CYCLE COMPLETE

11/15/2022 09:11 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____