

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 002556  
8:50 04/03/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:01  
PRESSURIZING 2:07  
STERILIZING 3:15  
136.4 C 224kPa 3:15  
136.3 C 224kPa 3:45  
136.2 C 223kPa 4:15  
136.2 C 222kPa 4:45  
136.4 C 225kPa 5:15  
136.4 C 224kPa 5:45  
136.3 C 224kPa 6:15  
136.1 C 221kPa 6:45  
136.1 C 221kPa 6:45  
VENTING 6:45  
AIR DRYING 7:34  
CYCLE COMPLETE 37:34  
-----

Digital Signature #  
7B401B057DDF18CA  
-----