
Midmark M11 - v1.0.5
Total Cycles: 3566

Name: RITTER2
Sterilizer ID: V2313228
04/26/2024 03:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.5 F 0.0
2:00 162.2 F 1.8
4:00 188.1 F 6.0
6:00 212.7 F 11.0
8:00 224.7 F 6.3
10:00 239.0 F 11.9
12:00 236.6 F 9.4
14:00 237.5 F 9.2
16:00 248.9 F 14.9
18:00 261.5 F 21.9
20:00 272.3 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 272.5 F 29.3
2:00 273.1 F 29.7
3:00 271.6 F 29.1
4:00 272.8 F 29.4

Min 271.6 F 29.0
Max 273.2 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.4 F 0.0
5:00 236.6 F 0.0
10:00 211.5 F 0.0
15:00 183.2 F 0.0
20:00 179.3 F 0.0
25:00 176.8 F 0.0
30:00 176.5 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 21:09
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:43
CYCLE COMPLETE

04/26/2024 04:21 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____