

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001459  
14:24 18/09/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:41  
PRESSURIZING 2:49  
STERILIZING 3:51  
136.0 C 220kPa 3:51  
136.4 C 225kPa 4:21  
136.4 C 225kPa 4:51  
136.2 C 223kPa 5:21  
135.9 C 220kPa 5:51  
136.4 C 225kPa 6:21  
136.4 C 225kPa 6:51  
136.3 C 224kPa 7:21  
136.3 C 224kPa 7:21  
VENTING 7:21  
AIR DRYING 8:09  
CYCLE COMPLETE 11:19

-----  
Digital Signature #  
6A18DD3B2A5DA497  
-----