
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001403
10:25 09/08/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:46
PRESSURIZING 5:22
STERILIZING 6:43
136.0 C 220kPa 6:43
135.8 C 218kPa 7:13
136.4 C 223kPa 7:43
136.4 C 224kPa 8:13
136.3 C 223kPa 8:43
136.2 C 222kPa 9:13
136.4 C 222kPa 9:43
136.5 C 224kPa 10:13
VENTING 10:13
AIR DRYING 11:17
CYCLE COMPLETE 11:18

Digital Signature #
C09618908983677B
