
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 008491
9:53 02/04/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:55
PRESSURIZING 2:58
STERILIZING 3:59
136.0 C 220kPa 3:59
136.4 C 224kPa 4:29
136.4 C 223kPa 4:59
136.4 C 222kPa 5:29
136.0 C 221kPa 5:59
136.4 C 221kPa 6:29
136.2 C 219kPa 6:59
136.3 C 224kPa 7:29
136.3 C 224kPa 7:29
VENTING 7:29
AIR DRYING 8:19
CYCLE COMPLETE 16:36

Digital Signature #
AF04258F7AAE4C23
