
Midmark M11 - v1.0.5
Total Cycles: 1944

Name: RITTER2
Sterilizer ID: V2313228
01/18/2023 10:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 206.6 F 0.0
2:00 193.1 F 3.9
4:00 209.3 F 9.7
6:00 226.3 F 7.0
8:00 237.8 F 11.9
10:00 236.7 F 9.8
12:00 236.6 F 8.8
14:00 250.7 F 16.0
16:00 262.8 F 23.2
18:00 271.6 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.2
1:00 272.4 F 29.3
2:00 273.0 F 29.7
3:00 272.3 F 29.8
4:00 272.3 F 29.2

Min 271.8 F 29.1
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.2 F 0.0
5:00 230.5 F 0.0
10:00 226.2 F 0.0
15:00 219.6 F 0.0
20:00 214.2 F 0.0
25:00 208.5 F 0.0
30:00 210.1 F 0.0
Drying Complete

FILLING: 2:00
HEATING: 18:46
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:56
CYCLE COMPLETE

01/18/2023 11:34 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____