

-----  
Midmark M11 - v1.0.0  
Total Cycles: 970

Name: RITTER 1  
Sterilizer ID: V2178198  
07/06/2021 02:01 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 211.8 F 0.0  
2:00 209.1 F 3.8  
4:00 214.8 F 8.5  
6:00 225.3 F 8.5  
8:00 231.7 F 9.2  
10:00 240.0 F 9.9  
12:00 242.8 F 12.5  
14:00 241.8 F 11.2  
16:00 254.9 F 18.0  
18:00 266.4 F 25.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.4  
1:00 272.5 F 28.9  
2:00 271.8 F 28.6  
3:00 272.4 F 28.9  
4:00 271.7 F 28.5  
  
Min 271.6 F 28.3  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.4 F 0.0  
5:00 233.9 F 0.0  
10:00 228.3 F 0.0  
15:00 228.9 F 0.0  
20:00 228.1 F 0.0  
25:00 226.0 F 0.0  
30:00 223.7 F 0.0  
Drying Complete

FILLING: 1:60  
HEATING: 19:33  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:58:47  
CYCLE COMPLETE

07/06/2021 02:59 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01