
Midmark M11 - v1.0.5
Total Cycles: 3844

Name: RITTER2
Sterilizer ID: V2313228
08/09/2024 05:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.2 F 0.0
2:00 72.3 F 0.0
4:00 86.5 F 0.0
6:00 108.7 F 0.1
8:00 148.3 F 0.4
10:00 170.1 F 3.2
12:00 187.2 F 6.3
14:00 201.6 F 9.5
16:00 213.7 F 12.7
18:00 220.4 F 6.4
20:00 230.6 F 10.3
22:00 233.9 F 9.1
24:00 241.1 F 12.2
26:00 239.9 F 10.8
28:00 251.9 F 16.8
30:00 262.6 F 23.0
32:00 272.6 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.0
1:00 272.6 F 29.6
2:00 272.2 F 29.7
3:00 271.6 F 29.0
4:00 272.3 F 29.4

Min 271.5 F 28.9
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.1 F 0.0
5:00 198.5 F 0.0
10:00 198.1 F 0.0
15:00 194.4 F 0.0
20:00 201.0 F 0.0
25:00 201.7 F 0.0
30:00 201.9 F 0.0

Drying Complete

FILLING: 0:50
HEATING: 32:14
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:10:35
CYCLE COMPLETE

08/09/2024 06:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____