

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5349

Name: RITTER 1  
Sterilizer ID: V2178198  
02/20/2025 01:10 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 173.6 F 0.0  
2:00 172.8 F 2.7  
4:00 182.5 F 6.0  
6:00 197.9 F 10.1  
8:00 209.5 F 7.1  
10:00 222.5 F 8.3  
12:00 234.3 F 12.7  
14:00 233.0 F 9.0  
16:00 243.7 F 13.0  
18:00 242.3 F 11.8  
20:00 252.6 F 16.9  
22:00 261.8 F 22.2  
24:00 270.6 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.6  
1:00 272.2 F 29.2  
2:00 272.8 F 29.2  
3:00 272.2 F 28.7  
4:00 271.7 F 28.7

Min 271.5 F 28.5  
Max 272.9 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.0 F 0.0  
5:00 191.9 F 0.0  
10:00 214.7 F 0.0  
15:00 202.8 F 0.0  
20:00 198.4 F 0.0  
25:00 202.1 F 0.0  
30:00 205.1 F 0.0  
Drying Complete

FILLING: 1:20  
HEATING: 25:00  
STERILIZING: 4:00  
VENTING: 3:52  
DRYING: 30:00  
TOTAL CYCLE: 01:04:13  
CYCLE COMPLETE

02/20/2025 02:14 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_