

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3975

Name: RITTER2  
Sterilizer ID: V2313228  
09/25/2024 03:55 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 99.4 F 0.0  
2:00 99.8 F 0.0  
4:00 107.4 F 0.1  
6:00 152.0 F 0.2  
8:00 180.2 F 2.5  
10:00 197.2 F 6.1  
12:00 210.7 F 9.8  
14:00 225.7 F 10.2  
16:00 229.9 F 9.4  
18:00 237.2 F 10.9  
20:00 240.9 F 12.1  
22:00 240.1 F 10.9  
24:00 252.6 F 17.2  
26:00 263.5 F 23.7  
28:00 272.3 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.5  
1:00 272.8 F 30.0  
2:00 271.6 F 29.5  
3:00 272.7 F 30.0  
4:00 271.7 F 29.4  
  
Min 271.6 F 29.3  
Max 272.9 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.3 F 0.0  
5:00 196.5 F 0.0  
10:00 213.5 F 0.0  
15:00 207.3 F 0.0  
20:00 197.4 F 0.0  
25:00 202.4 F 0.0  
30:00 201.8 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 29:02  
STERILIZING: 4:00  
VENTING: 3:41  
DRYING: 30:00  
TOTAL CYCLE: 01:07:40  
CYCLE COMPLETE

09/25/2024 05:02 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_