

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1354

Name: RITTER2  
Sterilizer ID: V2313228  
07/27/2022 09:39 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 126.9 F 0.0  
2:00 128.0 F 0.0  
4:00 133.0 F 0.1  
6:00 172.3 F 0.5  
8:00 196.3 F 4.0  
10:00 208.7 F 8.4  
12:00 222.6 F 12.9  
14:00 229.3 F 8.5  
16:00 240.7 F 13.5  
18:00 240.6 F 11.5  
20:00 239.4 F 10.4  
22:00 252.6 F 17.0  
24:00 264.0 F 23.8  
26:00 271.6 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.9  
1:00 272.7 F 29.7  
2:00 272.1 F 29.0  
3:00 271.7 F 29.2  
4:00 272.8 F 29.6

Min 271.4 F 28.7  
Max 273.0 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.4 F 0.0  
5:00 212.7 F 0.0  
10:00 225.9 F 0.0  
15:00 215.2 F 0.0  
20:00 214.8 F 0.0  
25:00 217.9 F 0.0  
30:00 218.1 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 26:08  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:04:08  
CYCLE COMPLETE

07/27/2022 10:43 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_