
Midmark M11 - v1.0.0
Total Cycles: 1684

Name: RITTER 1
Sterilizer ID: V2178198
03/01/2022 09:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 125.2 F 0.0
2:00 124.6 F 0.2
4:00 144.8 F 1.8
6:00 169.1 F 4.1
8:00 196.3 F 8.0
10:00 233.0 F 12.7
12:00 228.3 F 7.1
14:00 239.0 F 12.0
16:00 236.0 F 9.0
18:00 238.8 F 9.8
20:00 246.6 F 13.5
22:00 258.7 F 20.1
24:00 269.5 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.7 F 29.0
2:00 271.3 F 28.4
3:00 272.5 F 29.1
4:00 271.6 F 28.4

Min 271.2 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.9 F 0.0
5:00 213.0 F 0.0
10:00 198.5 F 0.0
15:00 199.2 F 0.0
20:00 202.2 F 0.0
25:00 208.8 F 0.0
30:00 221.9 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 24:59
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:03:04
CYCLE COMPLETE

03/01/2022 10:55 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____