
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006891
13:29 01/02/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:16
PRESSURIZING 2:21
STERILIZING 3:16
136.4 C 223kPa 3:16
136.0 C 221kPa 3:46
136.4 C 223kPa 4:16
136.2 C 223kPa 4:46
136.1 C 224kPa 5:16
136.4 C 224kPa 5:46
136.2 C 222kPa 6:16
136.4 C 224kPa 6:46
136.4 C 224kPa 6:46
VENTING 6:46
AIR DRYING 7:37
CYCLE COMPLETE 34:12

Digital Signature #
96994060591F0B2F
