
Midmark M11 - v1.0.5
Total Cycles: 4025

Name: RITTER2
Sterilizer ID: V2313228
10/11/2024 02:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 126.5 F 0.0
2:00 126.5 F 0.2
4:00 149.2 F 2.3
6:00 173.3 F 5.6
8:00 190.3 F 9.1
10:00 205.3 F 13.3
12:00 215.4 F 7.2
14:00 229.5 F 11.5
16:00 232.4 F 7.8
18:00 242.1 F 13.1
20:00 241.7 F 11.4
22:00 253.2 F 17.2
24:00 263.4 F 23.3
26:00 272.7 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.3
1:00 272.0 F 29.3
2:00 272.3 F 29.7
3:00 271.7 F 29.8
4:00 271.7 F 29.2

Min 271.1 F 29.2
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.9 F 0.0
5:00 228.8 F 0.0
10:00 183.6 F 0.0
15:00 174.9 F 0.0
20:00 172.1 F 0.0
25:00 170.8 F 0.0
30:00 170.2 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 27:02
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:05:39
CYCLE COMPLETE

10/11/2024 03:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____