
Midmark M11 - v1.0.5
Total Cycles: 2050

Name: RITTER2
Sterilizer ID: V2313228
02/14/2023 04:18 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.0 F 0.0
2:00 160.3 F 0.1
4:00 191.8 F 3.3
6:00 204.0 F 7.9
8:00 221.5 F 13.0
10:00 229.5 F 8.7
12:00 239.8 F 11.4
14:00 242.3 F 12.4
16:00 242.0 F 11.7
18:00 255.3 F 18.6
20:00 266.9 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.5 F 29.4
2:00 271.7 F 29.2
3:00 272.4 F 29.4
4:00 271.4 F 29.0

Min 271.4 F 28.8
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.2 F 0.0
5:00 208.8 F 0.0
10:00 231.2 F 0.0
15:00 229.1 F 0.0
20:00 226.7 F 0.0
25:00 224.5 F 0.0
30:00 224.5 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 21:26
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:42
CYCLE COMPLETE

02/14/2023 05:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____