
Midmark M11 - v1.0.5
Total Cycles: 1517

Name: RITTER2
Sterilizer ID: V2313228
09/21/2022 03:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.4 F 0.0
2:00 147.9 F 2.4
4:00 169.5 F 6.2
6:00 193.8 F 11.1
8:00 215.4 F 5.7
10:00 226.1 F 11.1
12:00 230.4 F 8.0
14:00 240.6 F 11.7
16:00 244.2 F 13.0
18:00 256.5 F 19.7
20:00 268.0 F 26.9
22:00 272.3 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.7
1:00 273.0 F 30.5
2:00 272.3 F 29.8
3:00 272.0 F 30.1
4:00 272.9 F 30.4

Min 271.6 F 29.6
Max 273.0 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.1 F 0.0
5:00 220.8 F 0.0
10:00 226.6 F 0.0
15:00 231.4 F 0.0
20:00 227.7 F 0.0
25:00 226.5 F 0.0
30:00 225.5 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 22:08
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:38
CYCLE COMPLETE

09/21/2022 04:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____