
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 003209
15:45 23/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:23
PRESSURIZING 2:26
STERILIZING 3:29
136.4 C 224kPa 3:29
136.4 C 225kPa 3:59
136.4 C 225kPa 4:29
136.4 C 223kPa 4:59
136.4 C 225kPa 5:29
136.4 C 225kPa 5:59
136.1 C 219kPa 6:29
136.2 C 221kPa 6:59
136.2 C 221kPa 6:59
VENTING 6:59
AIR DRYING 7:53
CYCLE COMPLETE 7:57

Digital Signature #
191A0BF03F14375E
