

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3751

Name: RITTER2  
Sterilizer ID: V2313228  
07/03/2024 09:49 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 160.1 F 0.0  
2:00 156.6 F 2.2  
4:00 165.0 F 5.7  
6:00 184.3 F 10.2  
8:00 213.9 F 5.2  
10:00 225.7 F 9.0  
12:00 235.9 F 10.8  
14:00 239.2 F 11.4  
16:00 237.4 F 9.8  
18:00 251.2 F 16.4  
20:00 262.8 F 23.2  
22:00 272.2 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 29.0  
1:00 272.2 F 29.7  
2:00 271.6 F 29.0  
3:00 271.5 F 29.6  
4:00 272.3 F 29.5

Min 271.0 F 28.9  
Max 272.5 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.4 F 0.0  
5:00 224.8 F 0.0  
10:00 196.2 F 0.0  
15:00 188.1 F 0.0  
20:00 187.2 F 0.0  
25:00 186.2 F 0.0  
30:00 184.4 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 22:30  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 01:00:53  
CYCLE COMPLETE

07/03/2024 10:49 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_