
Midmark M11 - v1.0.5
Total Cycles: 339

Name: RITTER2
Sterilizer ID: V2313228
09/08/2021 10:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.7 F 0.0
2:00 190.5 F 3.8
4:00 207.8 F 9.6
6:00 225.8 F 7.8
8:00 233.2 F 12.8
10:00 239.3 F 11.2
12:00 238.9 F 10.4
14:00 252.6 F 17.2
16:00 264.0 F 24.0
18:00 270.9 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 29.8
1:00 272.3 F 30.2
2:00 272.4 F 30.2
3:00 272.3 F 30.1
4:00 272.1 F 30.0

Min 271.9 F 29.7
Max 273.0 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.9 F 0.0
5:00 229.3 F 0.0
10:00 212.6 F 0.0
15:00 178.2 F 0.0
Dry Cycle Aborted

FILLING: 2:42
HEATING: 18:36
STERILIZING: 4:00
VENTING: 3:10
DRYING: 18:30
TOTAL CYCLE: 00:46:58

09/08/2021 11:10 AM
ACCEPTED: X
REJECTED:
Operator: U01