

-----  
Midmark M11 - v1.0.5  
Total Cycles: 693

Name: RITTER2  
Sterilizer ID: V2313228  
01/13/2022 09:18 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 138.8 F 0.0  
2:00 140.1 F 1.7  
4:00 158.0 F 4.6  
6:00 181.0 F 8.8  
8:00 203.2 F 9.4  
10:00 219.2 F 7.6  
12:00 232.9 F 12.8  
14:00 235.8 F 10.2  
16:00 235.7 F 8.6  
18:00 249.4 F 15.6  
20:00 261.4 F 22.5  
22:00 272.4 F 29.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.2  
1:00 272.7 F 29.5  
2:00 272.6 F 30.0  
3:00 272.0 F 29.1  
4:00 272.9 F 29.9

Min 271.7 F 29.1  
Max 273.0 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.6 F 0.0  
5:00 207.4 F 0.0  
10:00 216.9 F 0.0  
15:00 223.1 F 0.0  
20:00 229.4 F 0.0  
25:00 228.2 F 0.0  
30:00 215.8 F 0.0  
Drying Complete

FILLING: 1:20  
HEATING: 22:25  
STERILIZING: 4:00  
VENTING: 3:03  
DRYING: 30:00  
TOTAL CYCLE: 01:00:48  
CYCLE COMPLETE

01/13/2022 10:47 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02