
Midmark M11 - v1.0.0
Total Cycles: 2351

Name: RITTER 1
Sterilizer ID: V2178198
09/21/2022 01:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 212.8 F 0.0
2:00 207.0 F 4.5
4:00 213.5 F 9.2
6:00 226.4 F 8.1
8:00 231.0 F 9.1
10:00 239.1 F 10.6
12:00 240.8 F 11.6
14:00 238.5 F 9.8
16:00 251.5 F 16.3
18:00 262.6 F 22.7
20:00 272.7 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.7 F 29.3
2:00 272.0 F 28.6
3:00 271.7 F 28.7
4:00 272.7 F 29.3

Min 271.5 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.1 F 0.0
5:00 232.0 F 0.0
10:00 232.7 F 0.0
15:00 229.1 F 0.0
20:00 227.2 F 0.0
25:00 228.8 F 0.0
30:00 229.3 F 0.0
Drying Complete

FILLING: 1:45
HEATING: 20:23
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:19
CYCLE COMPLETE

09/21/2022 02:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____